

DEPARTMENT OF CHEMISTRY SACRED HEART COLLEGE (AUTONOMOUS) THEVARA, KOCHI – 682013

Tel: 0484 2870529, mail: chemistry@shcollege.ac.in, www.shcollege.ac.in

SHC/SERB-SURE 004872/2023-2024/Chilled Water Circulator

13.03.2024

QUOTATION NOTICE

As per the reference cited above, sealed competitive quotations are invited from dealers/ firms for supply and installation of **Chilled Water Circulator (1 No)**, in the Organic Chemistry Research Lab at Sacred Heart College (Autonomous), Thevara, subject to the following conditions.

SI No	Name of item	QTY
	Chilled Water Circulator	
1.	Specification	1
	1. Flow rate of 15LPM	
	2. Inbuilt cooling reservoir of 4 Liter and above	
	3. Temperature resolution of 1 °C	
	4. Cooling capacity of 750 W at 15 °C	
	5. Chiller temperature range: -10 °C to Ambient temperature	
	6. Should be supplied with silicone tubing with insulation	
	7. 1-Year Warranty	

The sealed envelope containing the quotations (Including Transportation & Installation Charges) should be superscribed with the reference number cited above and the same should reach Dr. Deepak D Prabhu, Principle Investigator, SERB-SURE project, Department of Chemistry, Sacred Heart College (Autonomous), Thevara, Kerala-682013 on or before 11.00 AM on 22.03.2024. Quotations will be opened at 02.00 PM on 25.03.2024 at the office.

Conditions:

- 1. Quoted price shall be inclusive of all taxes/freight/loading and unloading/installation.
- 2. Quoted rates should have at least six months validity period.
- 3. Warranty and supporting catalogues should be provided.
- 4. The right to accept or reject the quotations without assigning any reason rests entirely with the undersigned.
- 5. If the last date for receipt is declared a holiday, the next working day will be the last date for the purpose.

PI-SERB Project